Subject: Re: IDL and ActiveX CCD camera Posted by Mark Rivers on Sat, 03 Aug 2002 00:28:00 GMT

View Forum Message <> Reply to Message

Rick Towler <rtowler@u.washington.edu> wrote in message news:aieao7\$1urk\$1@nntp6.u.washington.edu...

- > If you have IDL5.5 I encourage you to try the new COM/ActiveX interface.
- > The only help you get from RSI is found in the "What's new in 55" pdf. Not
- > as much as one would like but enough to start doing some damage.

I have tried using the COM/ActiveX interface in IDL 5.5 to talk to CCD cameras from Roper Scientific (formerly Princeton Instruments) using their COM interface. Basically it does not work from IDL, while it works fine from Visual Basic.

I have discussed the problem with technical folks from RSI and they understand that the problem is on their end. Here is an extract from their response to my problem report:

- > ..., your example code pointed
- > out some shortcomings in the COM data management of IDL 5.5, with
- > respect to the Roper Scientific automation. Apparently, it's passing
- > some parameters to and from the automation using some techniques
- > that we hadn't anticipated. We'll attempt to address these in a future
- > release of IDL. In particular, the ::GetFrame method won't work properly
- > from IDL 5.5.

Mark Rivers