
Subject: Re: Constructiuon of SOP instance
Posted by [Robbie](#) on Tue, 30 Jan 2007 00:41:22 GMT
[View Forum Message](#) <> [Reply to Message](#)

The IDL Broker exhibits very strange behaviour if you don't generate both a new SOP instance *and* SeriesUID. If you are creating multiple secondary capture images, then use the same new SeriesUID for all output images, and use a separate SOP instance for each output image. This makes the Output Series unique from the Input Series.

You should use the UID generation function included on your E.Soft workstation. Otherwise, you can use the one I included in my GDLffDICOM package.

Robbie
