

---

Subject: Re: Constructioun of SOP instance

Posted by [Laurens](#) on Fri, 16 Feb 2007 13:56:10 GMT

[View Forum Message](#) <> [Reply to Message](#)

---

Robbie wrote:

> The IDL Broker exhibits very strange behaviour if you don't generate  
> both a new SOP instance \*and\* SeriesUID. If you are creating multiple  
> secondary capture images, then use the same new SeriesUID for all  
> output images, and use a separate SOP instance for each output image.  
> This makes the Output Series unique from the Input Series.

>

> You should use the UID generation function included on your E.Soft  
> workstation. Otherwise, you can use the one I included in my  
> GDLffDICOM package.

>

> Robbie

>

>

yeah going to use yours, the e.soft DLM didn't work out for me (it  
didn't like IDL 6.3 :())

Can ya tell me where i can find yours?

Thnx, Laurens

---