Subject: Re: Constructioon of SOP instance Posted by Laurens on Fri, 16 Feb 2007 13:56:10 GMT

View Forum Message <> Reply to Message

## Robbie wrote:

- > The IDL Broker exhibits very strange behaviour if you don't generate
- > both a new SOP instance \*and\* SeriesUID. If you are creating multiple
- > secondary capture images, then use the same new SeriesUID for all
- > output images, and use a separate SOP instance for each output image.
- > This makes the Output Series unique from the Input Series.

>

- > You should use the UID generation function included on your E.Soft
- > workstation. Otherwise, you can use the one I included in my
- > GDLffDICOM package.

> Robbie

>

yeah going to use yours, the e.soft DLM didn't work out for me (it didn't like IDL 6.3:()

Can ya tell me where i can find yours?

Thnx, Laurens