
Subject: Re: DICOM3_2_IDL

Posted by [paritosh](#) on Thu, 11 Jul 1996 07:00:00 GMT

[View Forum Message](#) <> [Reply to Message](#)

I read the MR (Siemens Vision) images by a brut-force method but it works fine.

Visions 1.5 T scanner images have a header of 6144 bytes.

```
point_lun,inunit,6144
```

```
read,data ; your 2D matrix typically 256x256
```

Note these are 16 bit/pixel images (where only 12 are actually used)

To extract imaging parameters TI, TE, TR and FOV, I pass the header through a parser (after converting each byte to a character) and it seems to work fine.

Paritosh
